

University of Dundee

Image guided Ablation

Xhang, Xinrui; Melzer, Andreas

Published in:
Scottish Medical Journal

DOI:
[10.1177/0036933020973637](https://doi.org/10.1177/0036933020973637)

Publication date:
2021

Licence:
CC BY-NC

Document Version
Publisher's PDF, also known as Version of record

[Link to publication in Discovery Research Portal](#)

Citation for published version (APA):
Xhang, X., & Melzer, A. (2021). Image guided Ablation. *Scottish Medical Journal*, 66(4), 175-177.
<https://doi.org/10.1177/0036933020973637>

General rights

Copyright and moral rights for the publications made accessible in Discovery Research Portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from Discovery Research Portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain.
- You may freely distribute the URL identifying the publication in the public portal.

Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Image guided ablation

Xinrui Zhang¹ and Andreas Melzer^{2,3} 

Scottish Medical Journal

2021, Vol. 66(4) 175–177

© The Author(s) 2020



Article reuse guidelines:

sagepub.com/journals-permissions

DOI: 10.1177/0036933020973637

journals.sagepub.com/home/scm



Ablation refers to the local application of optical, acoustic or electrical energy and cold as to induce irreversible cell injury, apoptosis and coagulative necrosis of tissues. By contrast to surgical excision, ablation is a minimally invasive treatment option, whereby the scarified tissue remains *in situ* and is being absorbed over several months and transformed to a scar. Clinical use of ablation encompasses the treatment of various tumors, including liver, lung, kidney, pancreatic, head and neck cancer and bone metastasis. Additionally, neurological disorders, particularly essential tremor and Parkinson's disease, can be treated by ablation of brain tissue or neuronal structures.¹ Relatively novel is the use electrical energy in a certain pattern that induces cell apoptosis without coagulation, referred to irreversible electroporation (IRE), see Rui Chen et al. in this issue of *SMJ*. In order to decide which kind of local cell destruction is useful and can be applied safely a thorough understanding of the underlying principles is essential. In addition, it is required to use an appropriate imaging technology to monitor and control the process of tissue destruction.

The process of energy-induced cell/tissue destruction consists of two phases through direct and indirect mechanisms. The direct damage of cells occurs rapidly after exposure of the target tissue to high temperature, alteration of the cell membrane, dysfunction of mitochondrial and inhibition of DNA replication. Changes of cell membrane fluidity and permeability are considered as the major cause of cell injury, leading to dysfunction of actin filaments and microtubules and impairment of facilitated diffusion across the cell membrane.² Mitochondria are affected by high temperature, increasing leakage of protons through the inner mitochondrial membrane and changing the ultrastructure in minutes.³ Besides the changes in cellular level, heat-induced denaturation of key replication enzymes DNA polymerase α and β , which is responsible for semiconservative DNA replication and DNA repair synthesis respectively, thereby inhibiting DNA replication.⁴ Denaturation of polymerase substrate chromatin, abnormal condensation of non-histone nuclear matrix proteins, disruption of RNA synthesis and the release of lysosomal enzymes are believed the mechanisms of heat-mediated reproductive cell death.

The indirect mechanism occurs via several mechanisms, including induction of apoptosis, the release of cytokines and stimulation of immune response. Apoptosis is increased in the peripheral zone of the central ablated lesion, which undergoes coagulative necrosis. Expression of essential apoptotic protein p53 was upregulated and bcl-2 was downregulated in human liver cancer tissues after ablation treatment.⁵ Release of pro-inflammatory cytokines such as interleukin-1 β (IL-1 β), IL-6, IL-8, IL-18 and tumor necrosis factor- α (TNF- α) increase in several hours to days after ablation maximize the anti-tumor response.⁶ Heat shock proteins (HSPs) are a large family of stress-induced proteins and play a key role in cell survival and development. Recently, HSP70 attracts interest in research and is concerned as a biomarker and potential anti-tumor targets. Preclinical and clinical evidences indicate that the upregulation of HSP70 expression may stimulate anti-tumor immunity by inducing various immunological processes and regulation of multiple pathways including stress-activated kinases pathway JNK, extracellular signal-regulated kinases ERK and cell cycle inhibitor p21.^{7,8}

Energy transmitted into target tumor lesion and induces local tumor destruction. According to the energy sources, the clinical techniques can be categorized as radiofrequency ablation (RFA), microwave ablation (MWA), laser ablation (LA), irreversible electroporation (IRE) and high-intensity focused ultrasound (HIFU/FUS). Each of the technologies shows strengths and weaknesses in the ablation of various tissues and organs.

Radiofrequency ablation (RFA) is the most accepted modality for tumor ablation using an electromagnetic energy source with frequencies at 200–1200 Hz

¹Scientist, ICCAS, Faculty of Medicine, University Leipzig, Germany

²Director, ICCAS, Faculty of Medicine, University Leipzig, Germany

³Foundation Director of IMSaT (Institute for Medical Science & Technology), IMSaT, University Dundee, UK

Corresponding author:

Andreas Melzer, ICCAS, University of Leipzig Faculty of Medicine, Semmelweisstrasse 14, Leipzig, 04107, Germany.

Email: a.melzer@dundee.ac.uk

1000-5000 Volt for generation of heat.⁹ One or more radiofrequency electrodes can be placed by open surgery, laparoscopic surgery or percutaneously, under the guidance of ultrasound, computed tomography (CT) or magnetic resonance imaging (MRI). RFA is considered a feasible option for treatment of unresectable tumors like primary hepato-cellular carcinoma other malignancies and metastatic lesions. RFA has proven to show a lower complication rate compared to surgical excision.¹⁰ However, due to the delivery of heat, the heating zone is limited to a few millimeters surrounding the active electrode with the risk that tissue may become dehydrated and desiccated during treatment. The heat-sink effect is a well-known limitation of RFA, describes the phenomena when heat is absorbed by blood flow, dissipating the thermal effect and reduce treatment efficiency.¹¹ Immune responses is induced after RFA, increasing tumor cell recognition and tumor-specific T cells, thereby supporting tumor-free survival of the patients.

The mechanism of MWA is similar to RFA, cause rotation of water molecules in cells, induce frictional heating and consequently cause coagulative necrosis. MWA applicator (antenna) delivers electromagnetic microwaves above 900 MHz and has emerged as an alternative method to RFA.⁹ In comparison to RFA, MWA generates a larger heating area (up to 2 cm away from antenna) and higher temperatures with shorter treatment duration, the tissue penetration is more uniform and heat-sink effect is less relevant.¹² From the biological perspective, WMA stimulates local inflammation and anti-tumor immunity to a lesser extend compared to other ablation techniques.¹³

LASER is another option for thermal ablation. It transports high energy light through a small flexible optical fiber (diameter 0.2–0.8 mm) into organs. The light absorbed by tissue is converted into heat. LA is superior to RFA and WMA in precision and treatment efficiency. Laser probes have the advantages of MR-compatibility, allowing for non-invasive MR-imaging guidance and temperature monitoring via MR-thermometry.¹⁴ Nevertheless, the limited ablation area (1–2 cm²) and accurate placement of laser applicators in tumor are the main challenges.

The term *irreversible electroporation* (IRE) describes the permeabilization of the lipid plasma membrane with high external direct and alternate current (DC/AC) pulsed electric fields.¹⁵ The initial formation of pore on cell membrane reversible, increasing the uptake of drug or plasmid DNA when sufficient level of energy is applied, the changes of membrane permeability are permanent and irreversible.¹⁶ Different from the other ablative techniques, IRE is a relatively new technique with non-thermal cell destruction mechanism, regards as an alternative for the tissues which

are not suitable for thermal ablation, especially in cardiology. Only cell membrane is affected by IRE, regeneration of blood vessels and bile ducts post-treatment is possible. The first study of IRE on human was published in 2011 for ablation of liver tumor, clinical trials of IRE ablation for treatment of prostate and pancreas cancer are undergoing.¹⁷ IRE also demonstrates good potential to overcome the thermal-related limitations and can be used to support chemotherapy, but the optimal physical parameters to obtain tissue-specific death remain unclear.

HIFU/FUS is an access free ablation technique and has got regulatory approval for ablation of uterine fibroids, prostate cancer, bone metastasis and treatment of essential tremor. Single or multi-element ultrasound transducer are positioned outside the body, ultrasound beams travel to the focus zone and generate heat in the target region. The temperature in the target can be elevated to 60–85°C in seconds, the tumor with large volume is treated by scanning the focal zone through the entire tumor in clinic.¹⁸ Treatment planning and real-time visualization can be realized with ultrasound-imaging or MRI. HIFU/FUS is feasible for both thermal and mechanical ablation because the application of acoustic waves at high energy can cause acoustic cavitation and temporarily generate pore on cell membrane (so-called sonoporation).¹⁹ The air or bones in the acoustic field may lead to scatter or absorption of acoustic interface and consequently result in incomplete destruction.

Image guidance and imaging follow-up post-ablation is essential for assessment of treatment outcomes. Various imaging modalities like B-mode sonography, contrast-enhanced US, MRI, PET-CT are available for rapid visualization and calculation of ablated area.^{20,21} The stiffness of ablated tissue is crucial and correlated with pathologic changes. Shear-wave elastography (SWE) is a reliable solution to display the stiffness of tissue after ablation with RFA,²² MWA, LA,²³ IRE²¹ and HIFU ablation²⁴ qualitatively and quantitatively. Sailev et al. reported the potential of targeted manipulation of apoptotic pathways by using high Intensity focused ultrasound for cancer treatment.²⁵ Cheng et al. describe in this issues of SJM the use of SWE to visualize IRE generated lesions in an VX2 tumour animal model. According to their findings SWE provides tissue stiffness information of different IRE ablation satges as a non-invasive method. SWE has already shown a great potential for detection prostatic lesion (Nabi et al.)²⁶ In combination with energetic tissue destruction it may allow to combine diagnosis and therapy in a “one stop shop”.

ORCID iD

Andreas Melzer  <https://orcid.org/0000-0002-7874-8946>

References

1. Chu KF and Dupuy DE. Thermal ablation of tumours: biological mechanisms and advances in therapy. *Nat Rev Cancer* 2014; 14: 199–208.
2. Fajardo LF, Egbert B, Marmor J, et al. Effects of hyperthermia in a malignant tumor. *Cancer* 1980; 45: 613–623.
3. Willis WT, Jackman MR, Bizeau ME, et al. Hyperthermia impairs liver mitochondrial function in vitro. *Am J Physiol Regul Integr Comp Physiol* 2000; 278: R1240–R1246.
4. Warters RL and Roti Roti JL. Hyperthermia and the cell nucleus. *Radiat Res* 1982; 92: 458–462.
5. Yi J, Wu L, Liu Z, et al. High-intensity focused ultrasound ablation induced apoptosis in human hepatocellular carcinoma. *Hepatogastroenterology* 2014; 61: 2336–2339.
6. Takaki H, Cornelis F, Kako Y, et al. Thermal ablation and immunomodulation: from preclinical experiments to clinical trials. *Diagn Interv Imaging* 2017; 98: 651–659.
7. Teng LS, Jin KT, Han N, et al. Radiofrequency ablation, heat shock protein 70 and potential anti-tumor immunity in hepatic and pancreatic cancers: a minireview. *Hepatobiliary Pancreat Dis Int* 2010; 9: 361–365.
8. Sherman MY and Gabai VL. Hsp70 in cancer: back to the future. *Oncogene* 2015; 34: 4153–4161.
9. Saldanha DF, Khiatani VL, Carrillo TC, et al. Current tumor ablation technologies: basic science and device review. *Semin Intervent Radiol* 2010; 27: 247–254.
10. Friedman M, Mikityansky I, Kam A, et al. Radiofrequency ablation of cancer. *Cardiovasc Intervent Radiol* 2004; 27: 427–434.
11. Zhu F and Rhim H. Thermal ablation for hepatocellular carcinoma: what's new in 2019. *Chin Clin Oncol* 2019; 8: 58.
12. Vogl TJ, Nour-Eldin NA, Hammerstingl RM, et al. Microwave ablation (MWA): basics, technique and results in primary and metastatic liver neoplasms – review article. *RoFo: Fortschritte auf dem gebiete der rontgenstrahlen und der.* 2017; 189: 1055–1066.
13. Dong BW, Zhang J, Liang P, et al. Sequential pathological and immunologic analysis of percutaneous microwave coagulation therapy of hepatocellular carcinoma. *Int J Hyperthermia* 2003; 19: 119–133.
14. Schena E, Saccomandi P and Fong Y. Laser ablation for cancer: past, present and future. *J Funct Biomater* 2017; 8: 19.
15. Wojtaszczyk A, Caluori G, Pesl M, et al. Irreversible electroporation ablation for atrial fibrillation. *J Cardiovasc Electrophysiol* 2018; 29: 643–651.
16. Aycock KN and Davalos RV. Irreversible electroporation: background, theory, and review of recent developments in clinical oncology. *Bioelectricity* 2019; 1: 214–234.
17. Thomson KR, Cheung W, Ellis SJ, et al. Investigation of the safety of irreversible electroporation in humans. *J Vasc Interv Radiol* 2011; 22: 611–621.
18. van den Bijgaart RJ, Eikelenboom DC, Hoogenboom M, et al. Thermal and mechanical high-intensity focused ultrasound: perspectives on tumor ablation, immune effects and combination strategies. *Cancer Immunol Immunother* 2017; 66: 247–258.
19. Tachibana K. Emerging technologies in therapeutic ultrasound: thermal ablation to gene delivery. *Hum Cell* 2004; 17: 7–15.
20. Romanato J, Menezes MR, Santos AO, et al. (18)F-FDG PET/CT performed immediately after percutaneous ablation to evaluate outcomes of the procedure: preliminary results. *Radiologia Brasileira* 2019; 52: 24–32.
21. Sugimoto K, Moriyasu F, Kobayashi Y, et al. Assessment of various types of US findings after irreversible electroporation in porcine liver: comparison with radiofrequency ablation. *J Vasc Interv Radiol* 2015; 26: 279–287.
22. Sugimoto K, Oshiro H, Ogawa S, et al. Radiologic-pathologic correlation of three-dimensional shear-wave elastographic findings in assessing the liver ablation volume after radiofrequency ablation. *World J Gastroenterol* 2014; 20: 11850–11855.
23. Gimenez ME, Davrieux CF, Saccomandi P, et al. Applications of elastography in ablation therapies: an animal ModelIn VivoStudy. *J Laparoendosc Adv S* 2020. DOI: 10.1089/lap.2020.0485.
24. Chen F, Li S, Wu LM, et al. Experimental study of applying shear wave elastography for examining effects of high intensity focused ultrasound ablation against subcutaneously implanted VX2 tumors in rabbits. *Biomed Res-India* 2017; 28: 7398–7403.
25. Saliev T, Feril LB, Jr, Nabi G, et al. Targeted manipulation of apoptotic pathways by using high intensity focused ultrasound in cancer treatment. *Cancer Lett* 2013; 338: 204–208.
26. Wei C, Li C, Szewczyk-Bieda M, et al. Performance characteristics of transrectal shear wave elastography imaging in the evaluation of clinically localized prostate cancer: a prospective study. *J Urol* 2018 ; 200: 549–558.